



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

5106
PERACETIC ACID
TEST STRIPS

LOT#: L-053040A

EXPIRATION DATE: 10/2011

Table with 6 columns: Date, Initials, Result, PASS, FAIL. Contains 7 rows of test data, all showing 'DARK BROWN' results and 'PASS' status.

NOTE: PLACE CHECKMARK IN BOX CORRESPONDING TO RESULT.
(PASS RESULTS INDICATED BY COLOR CHANGE TO GRAY/BLUE BLACK OR BROWN)
(FAIL RESULT IF PAD IS PREDOMINANTLY WHITE)

CONTROL DATA
Manufacturer: RPC
Part Number: MX-3212
Lot Number: 15081010

Sincerely,

Ted Williams
Quality Assurance / Regulatory Affairs Manager