



**Certificate of Conformance for Quality Assurance**  
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

**K100-0102**  
**TOTAL HARDNESS**  
**TEST STRIPS**

**LOT#: L-29071**

**EXPIRATION DATE: 08/2013**

	Date	Initials	Water Type	Result	Pass	Fail
1	08/10/11	TW	Calcium Reference	3 gpg	X	
2	08/10/11	TW	Calcium Reference	3 gpg	X	
3	08/10/11	TW	Calcium Reference	3 gpg	X	
4	08/10/11	TW	Calcium Reference	3 gpg	X	
5	08/10/11	TW	Distilled	0	X	
6	08/10/11	TW	Distilled	0	X	
7	08/10/11	TW	Distilled	0	X	

**NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT**  
 (FOR CALCIUM REFERENCE PASS RESULT INDICATED BY 3 GPG)  
 (FOR DISTILLED WATER PASS RESULT INDICATED BY 0 COLOR CHANGE)

<b>CONTROL DATA</b> Manufacturer: Hach Part Number: 22403-49 Lot Number: A1056 Expiration Date: 02/2012
---

Sincerely,

**Ted Williams**  
 Quality Assurance / Regulatory Affairs Manager