



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0107
RESIDUAL FORMALDEHYDE
TEST STRIP

LOT#: L-038010

EXPIRATION DATE: 07/2011

Table with 6 columns: ID, Date, Initials, Results, Pass, Fail. Contains 7 rows of test data.

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT
(PASS RESULT INDICATED BY COLOR DISTINCTLY DARKER THAN TRACE COLOR BLOCK)

Sincerely,

Ted Williams
Quality Assurance / Regulatory Affairs Manager