



Certificate of Conformance for Quality Assurance
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0113
BICARB pH
TEST STRIPS

LOT#: L-8137-0608

EXPIRATION DATE: 02/2010

	Date	Initials	7.0	7.5	8.0
1	06/30/2008	TW		X	
2	06/30/2008	TW		X	
3	06/30/2008	TW		X	
4	06/30/2008	TW		X	
5	06/30/2008	TW		X	
6	06/30/2008	TW		X	
7	06/30/2008	TW		X	

**NOTE: PLACE CHECKMARK IN BOX CORRESPONDING TO pH RESULT.
 PASSING TEST RESULT MUST BE BETWEEN 7.0 AND 8.0 pH**

CONTROL DATA	
MANUFACTURER:	HACH/ETS
PART NUMBER:	14079-95
LOT NUMBER:	A7346
EXPIRATION DATE:	12/2012

Sincerely,

Ted Williams
 Quality Assurance / Regulatory Affairs Manager