



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

5109
HISENSE CHLORINE/CHLORAMINE TEST STRIP

LOT#: L-048046

EXPIRATION DATE: 10/08

Table with 6 columns: date, initials, result, pass, fail. Rows 1-7 show test results with 'X' in the pass column and 'blue' in the result column. Rows 8-10 are empty.

NOTE: PLACE "X" IN BOX CORRESPONDING TO PASS OR FAIL RESULT
(PASS RESULTS INDICATED BY COLOR CHANGE TO
A DISTINCT BLUE COLOR FOR POSITIVE TEST
AND ZERO COLOR CHANGE FOR NEGATIVE TEST)

Sincerely,

Handwritten signature of Mike Lorenson

Mike Lorenson
Director, Quality Assurance & Regulatory Affairs

Form # 080106-2 A