



Certificate of Conformance for Quality Assurance
• *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0113
BICARB pH TEST STRIP

LOT#: L-7338-0108

EXPIRATION DATE: 09/09

	date	initials	7.0	7.5	8.0
1	01/23/08	TW		X	
2	01/23/08	TW		X	
3	01/23/08	TW		X	
4	01/23/08	TW		X	
5	01/23/08	TW		X	
6	01/23/08	TW		X	
7	01/23/08	TW		X	
8					
9					
10					

NOTE: PLACE "X" IN BOX CORRESPONDING TO pH RESULT.
(PASSING TEST RESULT MUST BE BETWEEN 7.0 AND 8.0 pH.)

Sincerely,

Ted Williams
Quality Assurance / Purchasing Manager

<p>CONTROL DATA Manufacturer: HACH/ETS Part Number: 14079-95 Lot Number: A6090 Expiration date: 03/2011</p>

Form # 022008-2 A