



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0118
ULTRA LOW TOTAL CHLORINE TEST STRIP

LOT#: L-T716-0807

EXPIRATION DATE: 07/09

Table with 6 columns: date, initials, result, pass, fail. Rows 1-7 show results of 0.1 ppm with 'X' in the pass column. Rows 8-10 are empty.

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT
(FOR PASS, RESULT MUST BE 0.1 PPM. ANYTHING OTHER THAN 0.1 PPM FAILS)

CONTROL DATA
Manufacturer: Chloramine T
Part Number: 276704E
Lot Number: K34920937-539

Sincerely,

Handwritten signature of Ted Williams

Ted Williams
Quality Assurance / Purchasing Manager

Form # 040408-1 A