



Certificate of Conformance for Quality Assurance

• *Quality Control Chart* •

#1000-5401

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0105

MICRO-X™ PERACETIC ACID POTENCY TEST STRIP

LOT#: L-05027-0507

EXPIRATION DATE: 02/09

	date	initials	result	pass	fail
1	05/31/07	KWB	dk brown	X	
2	05/31/07	KWB	dk brown	X	
3	05/31/07	KWB	dk brown	X	
4	05/31/07	KWB	dk brown	X	
5	05/31/07	KWB	dk brown	X	
6	05/31/07	KWB	dk brown	X	
7	05/31/07	KWB	dk brown	X	
8					
9					
10					

NOTE: PLACE "X" IN BOX CORRESPONDING TO PASS OR FAIL RESULT
(PASS RESULTS INDICATED BY COLOR CHANGE TO DARK BROWN ≥ 500 PPM)

Procedure 1000-5400 Section 4 Part 2.0 Specifications have been verified: X

Sincerely,

Mike Lorenson

Director, Quality Assurance & Regulatory Affairs

Form # 060907-4 A