



Reprocessing Products Corp (RPC)
 Rabrenco Scientific - Division of RPC

Certificate of Conformance for Quality Assurance
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0112
CHLORINE POTENCY TEST STRIP

LOT#: L-009056A

EXPIRATION DATE: 11/07

	date	initials	result	pass	fail
1	09/13/06	KWB	500 ppm	X	
2	09/13/06	KWB	500 ppm	X	
3	09/13/06	KWB	500 ppm	X	
4	09/13/06	KWB	500 ppm	X	
5	09/13/06	KWB	500 ppm	X	
6	09/13/06	KWB	500 ppm	X	
7	09/13/06	KWB	500 ppm	X	
8					
9					
10					

NOTE: PLACE "X" IN BOX CORRESPONDING TO PASS OR FAIL RESULT
 (PASS RESULTS INDICATED BY COLOR CHANGE TO BLUE = TO 500 PPM)

Sincerely,

Mike Lorenson
 Director, Quality Assurance & Regulatory Affairs

Form # 091406-2 A